



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Saga Environmental Science Inspection Association  
Environmental Business Division  
1-1-2 Hikari Saga-shi, Saga 840-0033***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system  
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Anion analysis in drinking water by ion chromatograph on the basis of Ministry of Health, Labour and Welfare Notification No.261  
(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President/Operations Manager

*Initial Accreditation Date:*

April 1, 2019

*Issue Date:*

April 1, 2019

*Expiration Date:*

May 31, 2021

*Accreditation No.:*

102780

*Certificate No.:*

L19-169

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Saga Environmental Science Inspection Association Environmental Business Division

21-1-2 Hikari Saga-shi, Saga 840-0033  
Contact Name: Yasutomo Kawagishi Phone: 0952-22-1651

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	REFERENCE VALUE
Chemical <sup>F</sup>	Drinking water	Nitrate Nitrogen and Nitrite Nitrogen	“SOP for Inspection based on Article 20 of Water Supply Act” (SKW3-analysis-002)  On the basis of: Simultaneous analysis by Ion chromatograph (Anion) Notification No. 261 Appended Table 13 and 16-2, Ministry of Health, Labour and Welfare	10 mg/L or less
		Nitrite Nitrogen		0.04 mg/L or less
		Fluorine and its compounds		Fluorine 0.8 mg/L or less
		Chloric acid		0.6 mg/L or less
		Chloride ion		200 mg/L or less

- The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this testing at its fixed location.